

SCHEDULE 4

(Descriptive markings and verification markings: Extract from Part 1 of OIML R 61 and additional marking)

Position

3.11.1 The filling instrument shall have a place for the application of verification marks. This place shall:

- be such that the part on which it is located cannot be removed from the filling instrument without damaging the marks,
- allow easy application of the mark without changing the metrological qualities of the filling instrument,
- be visible without the filling instrument having to be moved when it is in service.

Commencement Information

II Sch. 4 para. 3.11.1 in force at 17.7.2000, see [reg. 1\(1\)](#)

Changes to legislation:

There are currently no known outstanding effects for the The Weighing Equipment (Automatic Gravimetric Filling Instruments) Regulations 2000, Paragraph 3.11.1.