
Status: EU Directives are being published on this site to aid cross referencing from UK legislation. After IP completion day (31 December 2020 11pm) no further amendments will be applied to this version.

ANNEX

In Annex IV to Directive 2011/65/EU the following point 27 is added:

27. Lead in
- solders,
 - termination coatings of electrical and electronic components and printed circuit boards,
 - connections of electrical wires, shields and enclosed connectors,
- which are used in
- (a) magnetic fields within the sphere of 1 m radius around the isocentre of the magnet in medical magnetic resonance imaging equipment, including patient monitors designed to be used within this sphere, or
 - (b) magnetic fields within 1 m distance from the external surfaces of cyclotron magnets, magnets for beam transport and beam direction control applied for particle therapy.

Expires on 30 June 2020.