

SCHEDULE 3 **E+W+S**

Regulations 37, 38

PRESCRIBED LIMITS OF ERROR

1. Subject to paragraphs 2 and 3 below, the prescribed limits of error expressed in terms of “e”, the relevant verification scale interval of the device under test, shall be those set out in the Table below.

Table

<i>Classification of Machine</i>	<i>Limits of Error in Excess or Deficiency upon passing as fit for use for trade</i>		<i>Number of verification scale intervals for loads—see Note below</i>
Column 1	Column 2	Column 3	Column 4
Class I			
(first part)	0.5 e	1 e	not more than 50 000 e
(second part)	1 e	2 e	more than 50 000 e and not more than 200 000 e
(third part)	1.5 e	3 e	more than 200 000 e
Class II			
(first part)	0.5 e	1 e	not more than 5 000 e
(second part)	1 e	2 e	more than 5 000 e and not more than 20 000 e
(third part)	1.5 e	3 e	more than 20 000 e
Class III			
(first part)	0.5 e	1 e	not more than 500 e
(second part)	1 e	2 e	more than 500 e and not more than 2 000 e
(third part)	1.5 e	3 e	more than 2 000 e
Class IIII			
(first part)	0.5 e	1 e	not more than 50 e
(second part)	1 e	2 e	more than 50 e and not more than 200 e
(third part)	1.5 e	3 e	more than 200 e

Note

When testing a weight indicating or printing device, in a weighing mode in which tare devices other than those within paragraph (b) of the definition of “tare device” in regulation 2 above are operable, the ranges listed in column 4 of the Table shall be increased by the corresponding tare value.

Commencement Information

II Sch. 3 para. 1 in force at 2.5.2000, see **reg. 1(1)**

Changes to legislation: There are currently no known outstanding effects for the *The Weighing Equipment (Non-automatic Weighing Machines) Regulations 2000, SCHEDULE 3. (See end of Document for details)*

2. An indicating or printing device which, in any single weighing mode, has more than one verification scale interval, each of which relates exclusively to a particular part of the weighing range, shall have, for each particular part, those error allowances which would apply if the complete weighing range had the verification scale interval which relates to that particular part.

Commencement Information

I2 Sch. 3 para. 2 in force at 2.5.2000, see **reg. 1(1)**

3. In the case where—

- (a) a machine has had its stamp obliterated under regulation 43(2) of these Regulations because of an alteration or adjustment involving only the replacement or addition of a dial, chart or pointer, and
- (b) the purpose of the alteration or adjustment was to modify an imperial machine to indicate weight in metric units, and
- (c) (i) within the period of fifteen days following the making of the alteration or adjustment the requirements of paragraph 4 below were complied with, or
(ii) the period for complying with those requirements has not yet expired,

the appropriate prescribed limits of error upon the first retesting of such a machine with a view to its being passed as fit for use for trade shall be those set out in the Table in paragraph 1 above as being applicable in relation to the obliteration of stamps.

Commencement Information

I3 Sch. 3 para. 3 in force at 2.5.2000, see **reg. 1(1)**

4. The requirements referred to in paragraph 3(c)(i) above are that the chief inspector of weights and measures for the area in which the machine will first be used for trade after it has been altered or adjusted, or if the address at which it will be so used is not known, for the area in which the alteration or adjustment is carried out, is furnished by the person carrying out the alteration or adjustment with the following particulars, namely—

- (a) his name and address;
- (b) particulars by which the machine may be identified;
- (c) the name and address of the person who will first use the machine for trade after its alteration or adjustment and the address at which it will be so used or, if those particulars are not known, an address at which the machine will be available for inspection;
- (d) confirmation that the modification consisted only of the replacement or addition of a chart, dial or pointer.

Commencement Information

I4 Sch. 3 para. 4 in force at 2.5.2000, see **reg. 1(1)**

Changes to legislation:

There are currently no known outstanding effects for the The Weighing Equipment (Non-automatic Weighing Machines) Regulations 2000, SCHEDULE 3.