

ANNEX

DETERMINATION OF DICHLOROMETHANE AND 1,1,1-TRICHLOROETHANE IDENTIFICATION AND DETERMINATION OF NITROMETHANE

7. REPEATABILITY⁽¹⁾

For a nitromethane content of about 0,3 % (m/m), the difference between the results of two determinations carried out in parallel on the same sample should not exceed an absolute value of 0,03 % (m/m).

Status: EU Directives are being published on this site to aid cross referencing from UK legislation. After
IP completion day (31 December 2020 11pm) no further amendments will be applied to this version.

- (1) Norm ISO 5725.