

ANNEX I

THE OECD PRINCIPLES OF GOOD LABORATORY PRACTICE (GLP)

SECTION II

GOOD LABORATORY PRACTICE PRINCIPLES

1. Test facility organisation and personnel
 - 1.1. Test facility management's responsibilities
 1. Each test facility management should ensure that these principles of good laboratory practice are complied with, in its test facility.
 2. At a minimum it should:
 - (a) ensure that a statement exists which identifies the individual(s) within a test facility who fulfil the responsibilities of management as defined by these principles of good laboratory practice;
 - (b) ensure that a sufficient number of qualified personnel, appropriate facilities, equipment, and materials are available for the timely and proper conduct of the study;
 - (c) ensure the maintenance of a record of the qualifications, training, experience and job description for each professional and technical individual;
 - (d) ensure that personnel clearly understand the functions they are to perform and, where necessary, provide training for these functions;
 - (e) ensure that appropriate and technically valid standard operating procedures are established and followed, and approve all original and revised standard operating procedures;
 - (f) ensure that there is a quality assurance programme with designated personnel and assure that the quality assurance responsibility is being performed in accordance with these principles of good laboratory practice;
 - (g) ensure that for each study an individual with the appropriate qualifications, training, and experience is designated by the management as the study director before the study is initiated. Replacement of a study director should be done according to established procedures, and should be documented;
 - (h) ensure, in the event of a multisite study, that, if needed, a principal investigator is designated, who is appropriately trained, qualified and experienced to supervise the delegated phase(s) of the study. Replacement of a principal investigator should be done according to established procedures, and should be documented;
 - (i) ensure documented approval of the study plan by the study director;
 - (j) ensure that the study director has made the approval study plan available to the quality assurance personnel;
 - (k) ensure the maintenance of a historical file of all standard operating procedures;
 - (l) ensure that an individual is identified as responsible for the management of the archive(s);

Status: EU Directives are published on this site to aid cross referencing from UK legislation. Since IP completion day (31 December 2020 11.00 p.m.) no amendments have been applied to this version.

- (m) ensure the maintenance of a master schedule;
 - (n) ensure that test facility supplies meet requirements appropriate to their use in a study;
 - (o) ensure for a multisite study that clear lines of communication exist between the study director, principal investigator(s), the quality assurance programme(s) and study personnel;
 - (p) ensure that test and reference items are appropriately characterised;
 - (q) establish procedures to ensure that computerised systems are suitable for their intended purpose, and are validated, operated and maintained in accordance with these principles of good laboratory practice.
3. When a phase(s) of a study is conducted at a test site, test site management (if appointed) will have the responsibilities as defined above with the following exceptions: 1.1.2(g), (i), (j) and (o).

1.2. Study director's responsibilities

1. The study director is the single point of study control and has the responsibility for the overall conduct of the study and for its final report.
2. These responsibilities should include, but not be limited to, the following functions. The study director should:
 - (a) approve the study plan and any amendments to the study plan by dated signature;
 - (b) ensure that the quality assurance personnel have a copy of the study plan and any amendments in a timely manner and communicate effectively with the quality assurance personnel as required during the conduct of the study;
 - (c) ensure that study plans and amendments and standard operating procedures are available to study personnel;
 - (d) ensure that the study plan and the final report for a multisite study identify and define the role of any principal investigator(s) and any test facilities and test sites involved in the conduct of the study;
 - (e) ensure that the procedures specified in the study plan are followed, and assess and document the impact of any deviations from the study plan on the quality and integrity of the study, and take appropriate corrective action if necessary; acknowledge deviations from standard operating procedures during the conduct of the study;
 - (f) ensure that all raw data generated are fully documented and recorded;
 - (g) ensure that computerised systems used in the study have been validated;
 - (h) sign and date the final report to indicate acceptance of responsibility for the validity of the data and to indicate the extent to which the study complies with these principles of good laboratory practice;
 - (i) ensure that after completion (including termination) of the study, the study plan, the final report, raw data and supporting material are archived.

1.3. Principal investigator's responsibilities

The principal investigator will ensure that the delegated phases of the study are conducted in accordance with the applicable principles of good laboratory practice.

1.4. Study personnel's responsibilities

1. All personnel involved in the conduct of the study must be knowledgeable in those parts of the principles of good laboratory practice which are applicable to their involvement in the study.
2. Study personnel will have access to the study plan and appropriate standard operating procedures applicable to their involvement in the study. It is their responsibility to comply with the instructions given in these documents. Any deviation from these instructions should be documented and communicated directly to the study director, and/or if appropriate, the principal investigator(s).
3. All study personnel are responsible for recording raw data promptly and accurately and in compliance with these principles of good laboratory practice, and are responsible for the quality of their data.
4. Study personnel should exercise health precautions to minimise risk to themselves and to ensure the integrity of the study. They should communicate to the appropriate person any relevant known health or medical condition in order that they can be excluded from operations that may affect the study.

2. Quality assurance programme

2.1. General

1. The test facility should have a documented quality assurance programme to assure that studies performed are in compliance with these principles of good laboratory practice.
2. The quality assurance programme should be carried out by an individual or by individuals designated by and directly responsible to management and who are familiar with the test procedures.
3. This individual(s) should not be involved in the conduct of the study being assured.

2.2. Responsibilities of the quality assurance personnel

The responsibilities of the quality assurance personnel include, but are not limited to, the following functions. They should:

- (a) maintain copies of all approved study plans and standard operating procedures in use in the test facility and have access to an up-to-date copy of the master schedule;
- (b) verify that the study plan contains the information required for compliance with these principles of good laboratory practice. This verification should be documented;
- (c) conduct inspections to determine if all studies are conducted in accordance with these principles of good laboratory practice. Inspections should also determine that study plans and standard operating procedures have been made available to study personnel and are being followed.

Inspections can be of three types as specified by quality assurance programme standard operating procedures:

- study-based inspections,
- facility-based inspections,
- process-based inspections.

Records of such inspections should be retained;

Status: EU Directives are published on this site to aid cross referencing from UK legislation. Since IP completion day (31 December 2020 11.00 p.m.) no amendments have been applied to this version.

- (d) inspect the final reports to confirm that the methods, procedures, and observations are accurately and completely described, and that the reported results accurately and completely reflect the raw data of the studies;
- (e) promptly report any inspection results in writing to management and to the study director, and to the principal investigator(s) and the respective management, when applicable;
- (f) prepare and sign a statement, to be included with the final report, which specifies types of inspections and their dates, including the phase(s) of the study inspected, and the dates inspection results were reported to management and the study director and principal investigator(s), if applicable. This statement would also serve to confirm that the final report reflects the raw data.

3. Facilities

3.1. General

- 1. The test facility should be of suitable size, construction and location to meet the requirements of the study and to minimise disturbance that would interfere with the validity of the study.
- 2. The design of the test facility should provide an adequate degree of separation of the different activities to assure the proper conduct of each study.

3.2. Test system facilities

- 1. The test facility should have a sufficient number of rooms or areas to assure the isolation of test systems and the isolation of individual projects, involving substances or organisms known to be or suspected of being biohazardous.
- 2. Suitable rooms or areas should be available for the diagnosis, treatment and control of diseases, in order to ensure that there is no unacceptable degree of deterioration of test systems.
- 3. There should be storage rooms or areas as needed for supplies and equipment. Storage rooms or areas should be separated from rooms or areas housing the test systems and should provide adequate protection against infestation, contamination, and/or deterioration.

3.3. Facilities for handling test and reference items

- 1. To prevent contamination or mix-ups, there should be separate rooms or areas for receipt and storage of the test and reference items, and mixing of the test items with a vehicle.
- 2. Storage rooms or areas for the test items should be separate from rooms or areas containing the test systems. They should be adequate to preserve identity, concentration, purity, and stability, and ensure safe storage for hazardous substances.

3.4. Archive facilities

Archive facilities should be provided for the secure storage and retrieval of study plans, raw data, final reports, samples of test items and specimens. Archive design and archive conditions should protect contents from untimely deterioration.

3.5. Waste disposal

Handling and disposal of wastes should be carried out in such a way as not to jeopardise the integrity of studies. This includes provision for appropriate collection, storage and disposal facilities, and decontamination and transportation procedures.

4. Apparatus, material, and reagents
 1. Apparatus, including validated computerised systems, used for the generation, storage and retrieval of data, and for controlling environmental factors relevant to the study should be suitably located and of appropriate design and adequate capacity.
 2. Apparatus used in a study should be periodically inspected, cleaned, maintained, and calibrated according to standard operating procedures. Records of these activities should be maintained. Calibration should, where appropriate, be traceable to national or international standards of measurement.
 3. Apparatus and materials used in a study should not interfere adversely with the test systems.
 4. Chemicals, reagents, and solutions should be labelled to indicate identity (with concentration if appropriate), expiry date and specific storage instructions. Information concerning source, preparation date and stability should be available. The expiry date may be extended on the basis of documented evaluation or analysis.
5. Test systems
 - 5.1. Physical/chemical
 1. Apparatus used for the generation of physical/chemical data should be suitably located and of appropriate design and adequate capacity.
 2. The integrity of the physical/chemical test systems should be ensured.
 - 5.2. Biological
 1. Proper conditions should be established and maintained for the storage, housing, handling and care of biological test systems, in order to ensure the quality of the data.
 2. Newly received animal and plant test systems should be isolated until their health status has been evaluated. If any unusual mortality or morbidity occurs, this lot should not be used in studies and, when appropriate, should be humanely destroyed. At the experimental starting date of a study, test systems should be free of any disease or condition that might interfere with the purpose or conduct of the study. Test systems that become diseased or injured during the course of a study should be isolated and treated, if necessary to maintain the integrity of the study. Any diagnosis and treatment of any disease before or during a study should be recorded.
 3. Records of source, date of arrival, and arrival condition of test systems should be maintained.
 4. Biological test systems should be acclimatised to the test environment for an adequate period before the first administration/application of the test or reference item.
 5. All information needed to properly identify the test systems should appear on their housing or containers. Individual test systems that are to be removed from their housing or containers during the conduct of the study should bear appropriate identification, wherever possible.

Status: EU Directives are published on this site to aid cross referencing from UK legislation. Since IP completion day (31 December 2020 11.00 p.m.) no amendments have been applied to this version.

6. During use, housing or containers for test systems should be cleaned and sanitised at appropriate intervals. Any material that comes into contact with the test system should be free of contaminants at levels that would interfere with the study. Bedding for animals should be changed as required by sound husbandry practice. Use of pest control agents should be documented.
7. Test systems used in field studies should be located so as to avoid interference in the study from spray drift and from past usage of pesticides.
6. Test and reference items
 - 6.1. Receipt, handling, sampling and storage
 1. Records including test item and reference item characterisation, date of receipt, expiry date, quantities received and used in studies should be maintained.
 2. Handling, sampling, and storage procedures should be identified in order that the homogeneity and stability are assured to the degree possible and contamination or mix-up are precluded.
 3. Storage container(s) should carry identification information, expiry date, and specific storage instructions.
 - 6.2. Characterisation
 1. Each test and reference item should be appropriately identified (e.g. code, chemical abstracts service registry number (CAS number), name, biological parameters).
 2. For each study, the identity, including batch number, purity, composition, concentrations, or other characteristics to appropriately define each batch of the test or reference items should be known.
 3. In cases where the test item is supplied by the sponsor, there should be a mechanism, developed in cooperation between the sponsor and the test facility, to verify the identity of the test item subject to the study.
 4. The stability of test and reference items under storage and test conditions should be known for all studies.
 5. If the test item is administered or applied in a vehicle, the homogeneity, concentration and stability of the test item in that vehicle should be determined. For test items used in field studies (e.g. tank mixes), these may be determined through separate laboratory experiments.
 6. A sample for analytical purposes from each batch of test item should be retained for all studies except short-term studies.
 7. Standard operating procedures
 1. A test facility should have written standard operating procedures approved by test facility management that are intended to ensure the quality and integrity of the data generated by that test facility. Revisions to standard operating procedures should be approved by test facility management.
 2. Each separate test facility unit or area should have immediately available current standard operating procedures relevant to the activities being performed therein.

Status: EU Directives are published on this site to aid cross referencing from UK legislation. Since IP completion day (31 December 2020 11.00 p.m.) no amendments have been applied to this version.

Published text books, analytical methods, articles and manuals may be used as supplements to these standard operating procedures.

3. Deviations from standard operating procedures related to the study should be documented and should be acknowledged by the study director and the principal investigator(s), as applicable.
4. Standard operating procedures should be available for, but not be limited to, the following categories of test facility activities. The details given under each heading are to be considered as illustrative examples.
 1. Test and reference items
Receipt, identification, labelling, handling, sampling and storage.
 2. Apparatus, materials and reagents
 - (a) Apparatus:
use, maintenance, cleaning and calibration
 - (b) Computerised systems:
validation, operation, maintenance, security, change control and back-up
 - (c) Materials, reagents and solutions:
preparation and labelling.
 3. Record keeping, reporting, storage, and retrieval
Coding of studies, data collection, preparation of reports, indexing systems, handling of data, including the use of computerised systems.
 4. Test system (where appropriate)
 - (a) Room preparation and environmental room conditions for the test system.
 - (b) Procedures for receipt, transfer, proper placement, characterisation, identification and care of the test system.
 - (c) Test system preparation, observations and examinations, before, during and at the conclusion of the study.
 - (d) Handling of test system individuals found moribund or dead during the study.
 - (e) Collection, identification and handling of specimens including necropsy and histopathology.
 - (f) Siting and placement of test systems in test plots.
 5. Quality assurance procedures
Operation of Quality Assurance personnel in planning, scheduling, performing, documenting and reporting inspections.
 8. Performance of the study
 - 8.1. Study plan

Status: EU Directives are published on this site to aid cross referencing from UK legislation. Since IP completion day (31 December 2020 11.00 p.m.) no amendments have been applied to this version.

1. For each study, a written plan should exist prior to the initiation of the study. The study plan should be approved by dated signature of the study director and verified for GLP compliance by quality assurance personnel as specified in section II2(20) (b). The study plan should also be approved by the test facility management and the sponsor, if required by national regulation or legislation in the country where the study is being performed.
2.
 - (a) Amendments to the study plan should be justified and approved by dated signature of the study director and maintained with the study plan.
 - (b) Deviations from the study plan should be described, explained, acknowledged and dated in a timely fashion by the study director and/or principal investigator(s) and maintained with the study raw data.
3. For short-term studies, a general study plan accompanied by a study specific supplement may be used.

8.2. Content of the study plan

The study plan should contain, but not be limited to the following information:

1. Identification of the study, the test item and reference item
 - (a) A descriptive title
 - (b) A statement which reveals the nature and purpose of the study
 - (c) Identification of the test item by code or name (IUPAC; CAS number, biological parameters, etc.)
 - (d) The reference item to be used.
2. Information concerning the sponsor and the test facility
 - (a) Name and address of the sponsor
 - (b) Name and address of any test facilities and test sites involved
 - (c) Name and address of the study director
 - (d) Name and address of the principal investigator(s), and the phase(s) of the study delegated by the study director and under the responsibility of the principal investigator(s).
3. Dates
 - (a) The date of approval of the study plan by signature of the study director. The date of approval of the study plan by signature of the test facility management and sponsor if required by national regulation or legislation in the country where the study is being performed.
 - (b) The proposed experimental starting and completion dates.
4. Test methods
Reference to the OECD test guideline or other test guideline or method to be used.

5. Issues (where applicable)
 - (a) The justification for selection of the test system
 - (b) Characterisation of the test system, such as the species, strain, substrain, source of supply, number, body weight range, sex, age and other pertinent information
 - (c) The method of administration and the reason for its choice
 - (d) The dose levels and/or concentration(s), frequency, and duration of administration/application;
 - (e) Detailed information on the experimental design, including a description of the chronological procedure of the study, all methods, materials and conditions, type and frequency of analysis, measurements, observations and examinations to be performed, and statistical methods to be used (if any).
6. Records

A list of records to be retained.
- 8.3. Conduct of the study
 1. A unique identification should be given to each study. All items concerning this study should carry this identification. Specimens from the study should be identified to confirm their origin. Such identification should enable traceability, as appropriate for the specimen and study.
 2. The study should be conducted in accordance with the study plan.
 3. All data generated during the conduct of the study should be recorded directly, promptly, accurately, and legibly by the individual entering the data. These entries should be signed or initialled and dated.
 4. Any change in the raw data should be made so as not to obscure the previous entry, should indicate the reason for change and should be dated and signed or initialled by the individual making the change.
 5. Data generated as a direct computer input should be identified at the time of data input by the individual(s) responsible for direct data entries. Computerised system design should always provide for the retention of full audit trails to show all changes to the data without obscuring the original data. It should be possible to associate all changes to data with the persons having made those changes, for example, by use of timed and dated (electronic) signatures. Reason for changes should be given.
9. Reporting of study results
 - 9.1. General
 1. A final report should be prepared for each study. In the case of short-term studies, a standardised final report accompanied by a study specific extension may be prepared.
 2. Reports of principal investigators or scientists involved in the study should be signed and dated by them.
 3. The final report should be signed and dated by the study director to indicate acceptance of responsibility for the validity of the data. The extent of compliance with these principles of good laboratory practice should be indicated.

Status: EU Directives are published on this site to aid cross referencing from UK legislation. Since IP completion day (31 December 2020 11.00 p.m.) no amendments have been applied to this version.

4. Corrections and additions to a final report should be in the form of amendments. Amendments should clearly specify the reason for the corrections or additions and should be signed and dated by the study director.
5. Reformatting of the final report to comply with the submission requirements of a national registration or regulatory authority does not constitute a correction, addition or amendment to the final report.

9.2. Content of the final report

The final report should include, but not be limited to, the following information:

1. Identification of the study, the test item and reference item
 - (a) A descriptive title
 - (b) Identification of the test item by code or name (IUPAC, CAS number, biological parameters, etc.)
 - (c) Identification of the reference item by name
 - (d) Characterisation of the test item including purity, stability and homogeneity.
2. Information concerning the sponsor and the test facility
 - (a) Name and address of the sponsor
 - (b) Name and address of any test facilities and test sites involved
 - (c) Name and address of the study director
 - (d) Name and address of the principal investigator(s) and the phase(s) of the study delegated, if applicable
 - (e) Name and address of scientists having contributed reports to the final report.
3. Dates
Experimental starting and completion dates.
4. Statement
A quality assurance programme statement listing the types of inspections made and their dates, including the phase(s) inspected, and the dates any inspection results were reported to management and to the study director and principal investigator(s), if applicable. This statement would also serve to confirm that the final report reflects the raw data.
5. Description of materials and test methods
 - (a) Description of methods and materials used
 - (b) Reference to OECD test guideline or other test guideline or method.
6. Results
 - (a) A summary of results
 - (b) All information and data required by the study plan

- (c) A presentation of the results, including calculations and determinations of statistical significance
- (d) An evaluation and discussion of the results and, where appropriate, conclusions.

7. Storage

The location(s) where the study plan, samples of test and reference items, specimens, raw data and the final report are to be stored.

10. Storage and retention of records and materials

10.1. The following should be retained in the archives for the period specified by the appropriate authorities:

- (a) the study plan, raw data, samples of test and reference items, specimens, and the final report of each study;
- (b) records of all inspections performed by the quality assurance programme, as well as master schedules;
- (c) records of qualifications, training, experience and job descriptions of personnel;
- (d) records and reports of the maintenance and calibration of apparatus;
- (e) validation documentation for computerised systems;
- (f) the historical file of all standard operating procedures;
- (g) environmental monitoring records.

In the absence of a required retention period, the final disposition of any study materials should be documented. When samples of test and reference items and specimens are disposed of before the expiry of the required retention period for any reason, this should be justified and documented. Samples of test and reference items and specimens should be retained only as long as the quality of the preparation permits evaluation.

10.2. Material retained in the archives should be indexed so as to facilitate orderly storage and retrieval.

10.3. Only personnel authorised by management should have access to the archives. Movement of material in and out of the archives should be properly recorded.

10.4. If a test facility or an archive contracting facility goes out of business and has no legal successor, the archive should be transferred to the archives of the sponsor(s) of the study(s).