

ANNEX I

THE OECD PRINCIPLES OF GOOD LABORATORY PRACTICE (GLP)

SECTION II

GOOD LABORATORY PRACTICE PRINCIPLES

7. Standard operating procedures
 1. A test facility should have written standard operating procedures approved by test facility management that are intended to ensure the quality and integrity of the data generated by that test facility. Revisions to standard operating procedures should be approved by test facility management.
 2. Each separate test facility unit or area should have immediately available current standard operating procedures relevant to the activities being performed therein. Published text books, analytical methods, articles and manuals may be used as supplements to these standard operating procedures.
 3. Deviations from standard operating procedures related to the study should be documented and should be acknowledged by the study director and the principal investigator(s), as applicable.
 4. Standard operating procedures should be available for, but not be limited to, the following categories of test facility activities. The details given under each heading are to be considered as illustrative examples.
 1. Test and reference items
Receipt, identification, labelling, handling, sampling and storage.
 2. Apparatus, materials and reagents
 - (a) Apparatus:
use, maintenance, cleaning and calibration
 - (b) Computerised systems:
validation, operation, maintenance, security, change control and back-up
 - (c) Materials, reagents and solutions:
preparation and labelling.
 3. Record keeping, reporting, storage, and retrieval
Coding of studies, data collection, preparation of reports, indexing systems, handling of data, including the use of computerised systems.
 4. Test system (where appropriate)
 - (a) Room preparation and environmental room conditions for the test system.
 - (b) Procedures for receipt, transfer, proper placement, characterisation, identification and care of the test system.

Status: EU Directives are being published on this site to aid cross referencing from UK legislation. After IP completion day (31 December 2020 11pm) no further amendments will be applied to this version.

- (c) Test system preparation, observations and examinations, before, during and at the conclusion of the study.
 - (d) Handling of test system individuals found moribund or dead during the study.
 - (e) Collection, identification and handling of specimens including necropsy and histopathology.
 - (f) Siting and placement of test systems in test plots.
5. Quality assurance procedures
- Operation of Quality Assurance personnel in planning, scheduling, performing, documenting and reporting inspections.