

---

**Status:** EU Directives are being published on this site to aid cross referencing from UK legislation. After IP completion day (31 December 2020 11pm) no further amendments will be applied to this version.

---

## ANNEX

In Annex IV to Directive 2011/65/EU the following point 23 is added:

23. Lead as an alloying element for bearings and wear surfaces in medical equipment exposed to ionising radiation. Expires on 30 June 2021.