

## ANNEX XI

## CHAPTER II

**Length measuring instruments****MPEs**3. *Instrument*

TABLE 2

<b>Accuracy class</b>	<b>MPE</b>
I	0,125 %, but not less than 0,005 $L_m$
II	0,25 %, but not less than 0,01 $L_m$
III	0,5 %, but not less than 0,02 $L_m$

Where  $L_m$  is the minimum measurable length, that is to say the smallest length specified by the manufacturer for which the instrument is intended to be used.

The true length value of the different types of materials shall be measured using suitable instruments (e.g. tapes of length). Thereby, the material which is going to be measured shall be laid out on a suitable underlay (e.g. a suitable table) straight and unstretched.