

ANNEX

In Annex IV to Directive 2011/65/EU the following point 27 is added:

27. Lead in
- solders,
 - termination coatings of electrical and electronic components and printed circuit boards,
 - connections of electrical wires, shields and enclosed connectors,
- which are used in
- (a) magnetic fields within the sphere of 1 m radius around the isocentre of the magnet in medical magnetic resonance imaging equipment, including patient monitors designed to be used within this sphere, or
 - (b) magnetic fields within 1 m distance from the external surfaces of cyclotron magnets, magnets for beam transport and beam direction control applied for particle therapy.

Expires on 30 June 2020.