

ANNEX

In Annex III to Directive 2011/65/EU, point 7(c)-I is replaced by the following:

7(c)-I	Electrical and electronic components containing lead in a glass or ceramic other than dielectric ceramic in capacitors, e.g. piezoelectronic devices, or in a glass or ceramic matrix compound	Applies to categories 1-7 and 10 (except applications covered under point 34) and expires on 21 July 2021. For categories 8 and 9 other than <i>in vitro</i> diagnostic medical devices and industrial monitoring and control instruments expires on 21 July 2021. For category 8 <i>in vitro</i> diagnostic medical devices expires on 21 July 2023. For category 9 industrial monitoring and control instruments, and for category 11 expires on 21 July 2024.
--------	--	--