

---

**Status:** EU Directives are being published on this site to aid cross referencing from UK legislation. After IP completion day (31 December 2020 11pm) no further amendments will be applied to this version.

---

## ANNEX

In Annex III to Directive 2011/65/EU, point 7(a) is replaced by the following:

7(a)	Lead in high melting temperature type solders (i.e. lead-based alloys containing 85 % by weight or more lead)	Applies to categories 1-7 and 10 (except applications covered by point 24 of this Annex) and expires on 21 July 2021. For categories 8 and 9 other than in vitro diagnostic medical devices and industrial monitoring and control instruments expires on 21 July 2021. For category 8 in vitro diagnostic medical devices expires on 21 July 2023. For category 9 industrial monitoring and control instruments, and for category 11 expires on 21 July 2024.
------	---	--