

ANNEX

In Annex III, entry 29 is replaced by the following:

‘29	Lead bound in crystal glass as defined in Annex I (Categories 1, 2, 3 and 4) of Council Directive 69/493/EEC ^a	Expires on: — 21 July 2021 for categories 1-7 and 10; — 21 July 2021 for categories 8 and 9 other than in vitro diagnostic medical devices and industrial monitoring and control instruments; — 21 July 2023 for category 8 in vitro diagnostic medical devices; — 21 July 2024 for category 9 industrial monitoring and control instruments, and for category 11.
-----	---	--

^a Council Directive 69/493/EEC of 15 December 1969 on the approximation of the laws of the Member States relating to crystal glass (OJ L 326, 29.12.1969, p. 36).’