

Status: EU Directives are being published on this site to aid cross referencing from UK legislation. After IP completion day (31 December 2020 11pm) no further amendments will be applied to this version.

ANNEX

In Annex III, entry 18(b) is replaced by the following:

18(b)	Lead as activator in the fluorescent powder (1 % lead by weight or less) of discharge lamps when used as sun tanning lamps containing phosphors such as BSP (BaSi2O5:Pb)	Expires on: — 21 July 2021 for categories 1-7 and 10; — 21 July 2021 for categories 8 and 9 other than in vitro diagnostic medical devices and industrial monitoring and control instruments; — 21 July 2023 for category 8 in vitro diagnostic medical devices; — 21 July 2024 for category 9 industrial monitoring and control instruments, and for category 11.
18(b)-I	Lead as activator in the fluorescent powder (1 % lead by weight or less) of discharge lamps containing phosphors such as BSP (BaSi2O5:Pb) when used in medical phototherapy equipment	Applies to categories 5 and 8, excluding applications covered by entry 34 of Annex IV, and expires on 21 July 2021.