
STATUTORY RULES OF NORTHERN IRELAND

1998 No. 48

**Measuring Equipment (Capacity Measures)
Regulations (Northern Ireland) 1998**

Part IV

Supplementary Provisions

Prescribed limits of error

15. The prescribed limits of error—

- (a) for dispensing measures for pharmaceutical purposes, shall be those set out in BS 604: 1982 or BS 1922: 1987, as appropriate; and
- (b) for all other capacity measures shall—
 - (i) in the case of line measures, be the amounts set out in column 3 of Schedule 2 in excess and in deficiency; and
 - (ii) in the case of brim measures, be the amounts set out in column 4 of that Schedule in excess only.