
STATUTORY INSTRUMENTS

1963 No. 1710

The Weights and Measures Regulations 1963

PART III

LIQUID CAPACITY MEASURES

Testing

28.—(1) Liquid capacity measures of maximum purported values between 8 gallons and one-sixth of a gill inclusive, shall be tested by transferring water from the local or working standard into the measure under test.

(2) Measures—

- (a) with a lip or rim, shall be tested to the bottom of the lip or rim;
- (b) on which the purported value is defined by a line, shall be tested to the top of the line by taking the level of the water at the bottom of the meniscus.