

SCHEDULE 1

PRESCRIBED LIMITS OF ERROR UPON THE TESTING OF MEASURES AND WEIGHTS

PART I

LINEAR MEASURES

1. The prescribed limits of error for linear measures shall be—
 - (a) in relation to the passing of any such measure as fit for use for trade, the appropriate amount specified in paragraph 2 or 3, as the case may be, of this Part of this Schedule;
 - (b) in relation to the obliteration of the stamp on any such measure, an amount equal to four times the appropriate amount specified in paragraph 2 or 3, as the case may be, of this Part of this Schedule.

2. *Imperial system*

<i>End measures</i> <i>Measures made of metal: purported value</i>	<i>Error in excess</i>	<i>Line measures</i>		
		<i>Error in deficiency</i>	<i>Error in excess</i>	<i>Error in deficiency</i>
inches	inches	inches	inches	
Under 1 foot	0.01	0.01	0.005	0.002
1 foot to 1 yard inclusive	0.03	0.015	0.02	0.01
Above 1 yard and under 10 feet	—	—	0.05	0.05
10 feet and under 50 feet	—	—	0.2	0.2
50 feet to 100 feet inclusive	—	—	0.3	0.3

In the case of measures made of material other than metal, the foregoing amounts of error shall be increased to double the said amounts.

3. *Metric system*

<i>End measures</i> <i>Measures made of metal: purported value</i>	<i>Error in excess</i>	<i>Line measures</i>		
		<i>Error in deficiency</i>	<i>Error in excess</i>	<i>Error in deficiency</i>
millimetres	millimetres	millimetres	millimetres	
1 millimetre	0.25	0.025	0.05	0.025
1 centimetre	0.2	0.1	0.1	0.05
1 decimetre	0.5	0.25	0.2	0.1
1 metre	1	0.5	0.5	0.5
2 and 3 metres	2	1	1	1
10 metres	—	—	5	5
20 metres	—	—	7.5	7.5

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

In the case of measures made of material other than metal, the foregoing amounts of error shall be increased to double the said amounts.

PART II

CAPACITY MEASURES

1.—(1) The prescribed limits of error for capacity measures in relation to the passing of any such measure as fit for use for trade shall be the appropriate amount specified in paragraph 2 or 3, as the case may be, of this Part of this Schedule.

(2) The prescribed limits of error for capacity measures in relation to the obliteration of the stamp on any such measure shall be—

- (a) in the case of a capacity measure other than an apothecaries measure—
 - (i) if the error found on testing is in deficiency, an amount equal to half the corresponding amount prescribed in relation to the passing of such measure as fit for use for trade;
 - (ii) if the error so found is in excess, an amount equal to the corresponding amount prescribed in relation to the passing of such measure as fit for use for trade;
- (b) in the case of an apothecaries measure (and whether the error found on testing is in deficiency or excess), an amount equal to the corresponding amount prescribed in relation to the passing of such measure as fit for use for trade.

2. *Imperial system*

- (a) *Liquid measures, other than apothecaries measures*

<i>Purported value</i>	<i>Error in excess only</i>
# gill	½ fluid drachm
# gill	½ fluid drachm
¼ gill	½ fluid drachm
⅓ gill	1 fluid drachm
# gill	1 fluid drachm
½ gill	1 fluid drachm
4 fluid ounces	1 fluid drachm
1 gill	2 fluid drachms
6 fluid ounces	2 fluid drachms
⅓ pint	2 fluid drachms
8 fluid ounces	3 fluid drachms
½ pint	3 fluid drachms
1 pint	4 fluid drachms
1 quart	1 fluid ounce
½ gallon	1 fluid ounce
1 to 3 gallons inclusive	2 fluid ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value</i>	<i>Error in excess only</i>
4 to 7 gallons inclusive	3 fluid ounces
8 to 19 gallons inclusive	5 fluid ounces
20 to 32 gallons inclusive	10 fluid ounces
33 to 64 gallons inclusive	15 fluid ounces

- (i) In the case of conical-shaped measures made of metal, the foregoing amounts of error (as tabulated) shall be decreased to half the said amounts.
- (ii) In the case of milk churns of purported values of 4 to 32 gallons inclusive, the foregoing amounts of error (as tabulated) shall be increased to double the said amounts.
- (iii) In the case of measures made of enamelled-metal, glass or earthenware where the purported value is defined by the brim, and of a purported value exceeding half a pint, the foregoing amounts of error (as tabulated) shall be increased to double the said amounts; and of a purported value of half a pint, the prescribed limit of error shall be half a fluid ounce in excess only.
- (iv) In the case of subdivided measures, the error at any graduation shall not exceed that specified for a measure of equivalent purported value.

(b) *Apothecaries measures*

<i>Approximate internal diameter of measure at the graduation tested</i>	<i>Error in excess or in deficiency</i>
Inches	Minims
$\frac{1}{2}$	$\frac{1}{2}$
$\frac{5}{8}$	1
$\frac{3}{4}$	2
$\frac{7}{8}$	3
1	4
1 $\frac{1}{4}$	6
1 $\frac{1}{2}$	7
1 $\frac{3}{4}$	9
2	11
2 $\frac{1}{2}$	14
3	18
3 $\frac{1}{2}$	21
4	25

In the case of graduated measures made of glass in the form of burettes, the foregoing amounts of error shall be decreased to half the said amounts.

(c) *Dry measures*

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value</i>	<i>Error in excess only</i>
½ pint	5 fluid drachms
1 pint	¼ gill
1 quart	½ gill
½ gallon	1 gill
1 gallon	1 ½ gills
1 peck (2 gallons)	1 ½ gills
½ bushel (4 gallons)	½ pint
1 bushel (8 gallons)	½ pint

3. Metric System

(a) *Liquid measures made of metal, other than apothecaries measures.*

<i>Purported value</i>	<i>Error in excess only</i>
millilitre	
1 millilitres	0.05
2 millilitres	0.1
5 millilitres	0.25
10 millilitres	0.5
20 millilitres	1
25 millilitres	1
50 millilitres	2
100 millilitres	2
200 millilitres	5
250 millilitres	5
500 millilitres	10
1 litre	15
2 litres	25
2 ½ litres	25
5 litres	50
10 litres	75
20 litres	100

In the case of subdivided measures, the error at any graduation shall not exceed that specified for a measure of equivalent purported value.

(b) *Liquid measures made of earthenware, glass or enamelled-metal, and measures made of other materials approved by the Board, other than apothecaries measures:—*

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value</i>	<i>Error in excess only</i>
millilitres	
200 millilitres	10
250 millilitres	10
500 millilitres	25
1 litre	50
2 litres	100
2½ litre	100
5 litre	200

In the case of subdivided measures, the error at any graduation shall not exceed that specified for a measure of equivalent purported value.

(c) *Apothecaries measures*

<i>Approximate internal diameter of measure in millimetres at the graduation tested</i>	<i>Error in excess or in deficiency</i>
millimetres	millilitres
10	0.05
20	0.15
30	0.3
40	0.4
50	0.6
60	0.6
70	0.8
80	0.8
90	1
100	1

In the case of graduated measures made of glass in the form of burettes, the foregoing amounts of error shall be decreased to half the said amounts.

(d) *Dry measures*

<i>Purported value</i>	<i>Error in excess only</i>
millilitres	
200 millilitres	10
500 millilitres	25
1 litre	50

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value</i>	<i>Error in excess only</i>
2 litre	100
2½ litres	100
5 litres	150
10 litres	250
20 litres	300

PART III

WEIGHTS

1. The prescribed limits of error for weights (in relation both to passing the same as fit for use for trade and obliterating the stamp thereon) shall be those specified in paragraphs 2 and 3 of this Part of this Schedule.

2. Imperial system

(a) Avoirdupois weights:—

<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>		
<i>Purported value</i>	<i>Weights made of iron</i>	<i>Weights not made of iron</i>
½ dram	—	0.5 grain
1 dram	—	0.5 grain
2 drams	—	0.5 grain
4 drams	—	0.5 grain
8 drams	—	0.5 grain
1 ounce	—	1 grain
2 ounces	—	1 grain
4 ounces	4 grains	2 grains
8 ounces	4 grains	2 grains
1 pound	4 grains	2 grains
2 pounds	6 grains	3 grains
4 pounds	10 grains	5 grains
5 pounds	10 grains	5 grains
7 pounds	10 grains	5 grains
10 pounds	16 grains	8 grains
14 pounds	20 grains	10 grains
20 pounds	20 grains	10 grains
28 pounds	30 grains	15 grains

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>		
<i>Purported value</i>	<i>Weights made of iron</i>	<i>Weights not made of iron</i>
50 pounds	40 grains	20 grains
56 pounds	50 grains	25 grains

(b) *Grain weights:—*

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
0.01 grain	0.001 grain
0.02 grain	0.002 grain
0.03 grain	0.003 grain
0.05 grain	0.003 grain
0.1 grain	0.005 grain
0.2 grain	0.01 grain
0.3 grain	0.02 grain
0.5 grain	0.03 grain
1 grain	0.03 grain
2 grains	0.03 grain
3 grains	0.03 grain
5 grains	0.03 grain
10 grains	0.03 grain
20 grains	0.05 grain
30 grains	0.05 grain
50 grains	0.05 grain
100 grains	0.05 grain

(c) *Troy weights:—*

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
0.001 ounce troy	0.03 grain
0.002 ounce troy	0.03 grain
0.003 ounce troy	0.03 grain
0.004 ounce troy	0.03 grain
0.005 ounce troy	0.03 grain
0.01 ounce troy	0.03 grain

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
0.02 ounce troy	0.03 grain
0.025 ounce troy	0.03 grain
0.03 ounce troy	0.03 grain
0.04 ounce troy	0.05 grain
0.05 ounce troy	0.05 grain
0.1 ounce troy	0.05 grain
0.2 ounce troy	0.05 grain
0.3 ounce troy	0.1 grain
0.4 ounce troy	0.1 grain
0.5 ounce troy	0.1 grain
1 ounces troy	0.2 grain
2 ounces troy	0.2 grain
3 ounces troy	0.3 grain
4 ounces troy	0.3 grain
5 ounces troy	0.3 grain
10 ounces troy	0.5 grain
20 ounces troy	1 grains
30 ounces troy	1 grains
40 ounces troy	2 grains
50 ounces troy	2 grains
100 ounces troy	3 grains
200 ounces troy	3 grains
300 ounces troy	4 grains
400 ounces troy	4 grains
500 ounces troy	4 grains

(d) *Apothecaries weights:—*

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
4 grains	0.03 grain
6 grains	0.03 grain
½ scruple	0.03 grain
1 scruple	0.05 grain

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
1 ½ scruples	0.05 grain
2 scruples	0.05 grain
1 drachm	0.05 grain
2 drachms	0.1 grain
4 drachms	0.1 grain
1 ounce apothecaries	0.2 grain
2 ounces apothecaries	0.2 grain
4 ounces apothecaries	0.3 grain
6 ounces apothecaries	0.3 grain
8 ounces apothecaries	0.5 grain
10 ounces apothecaries	0.5 grain

(e) Pennyweights—

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
1 pennyweight	0.05 grain
2 pennyweights	0.05 grain
5 pennyweights	0.05 grain
10 pennyweights	0.1 grain

Metric system

(a) (a) Metric weights; other than carat (metric) weights:—

<i>Purported value</i> <i>Weights made of iron</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i> <i>Weights not made of iron</i>
1 milligramme	— 0.1 milligramme
2 milligrammes	— 0.2 milligrammes
5 milligrammes	— 0.2 milligrammes
10 milligrammes	— 0.5 milligrammes
20 milligrammes	— 1 milligramme
50 milligrammes	— 2 milligrammes
100 milligrammes	— 2 milligrammes

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value Weights made of iron</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>	
	<i>Weights not made of iron</i>	
200 milligrammes	—	2 milligrammes
500 milligrammes	—	2 milligrammes
1 gramme	—	5 milligrammes
2 grammes	—	5 milligrammes
5 grammes	—	5 milligrammes
10 grammes	—	10 milligrammes
20 grammes	—	15 milligrammes
50 grammes	—	15 milligrammes
100 grammes	40 milligrammes	20 milligrammes
200 grammes	100 milligrammes	50 milligrammes
500 grammes	200 milligrammes	100 milligrammes
1 kilogramme	400 milligrammes	200 milligrammes
2 kilogrammes	600 milligrammes	300 milligrammes
5 kilogrammes	1 gramme	500 milligrammes
10 kilogrammes	2 grammes	1 gramme
20 kilogrammes	3 grammes	1.5 grammes

(b) *Carat (metric) weights:—*

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
0.01 carat (metric)	0.2 milligramme
0.02 carat (metric)	0.2 milligramme
0.05 carat (metric)	0.2 milligramme
0.1 carat (metric)	0.2 milligramme
0.2 carat (metric)	0.5 milligramme
0.25 carat (metric)	0.5 milligramme
0.5 carat (metric)	0.5 milligramme
1 carat (metric)	1 milligramme
2 carats (metric)	1 milligramme
5 carats (metric)	1 milligramme
10 carats (metric)	1 milligramme

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Purported value</i>	<i>Error in excess only (or, in relation to the obliteration of stamps, error in excess or deficiency)</i>
20 carats (metric)	1 milligramme
50 carats (metric)	2 milligrammes
100 carats (metric)	2 milligrammes
200 carats (metric)	5 milligrammes
500 carats (metric)	5 milligrammes

SCHEDULE 2

PART I

PRESCRIBED LIMITS OF ERROR ON THE TESTING OF WEIGHING INSTRUMENTS

1. Subject to paragraphs 2 and 3 of this Part of this Schedule, the prescribed limits of error for weighing instruments shall be those specified in Parts II to XI of this Schedule:

Provided that in the case of any weighing instrument of a capacity not so specified, the prescribed limits of error shall be the amounts proportionate to those so specified for an instrument of the same type, class or description.

2. In the case of any weighing instrument which weighs in units of the metric system and for which no limits of error are specified in terms of those units, the prescribed limits of error shall be the amounts in terms of metric units equivalent to those specified in terms of imperial units in the relevant Part of this Schedule with respect to an instrument of the same capacity, type, class or description.

3. In the case of any weighing instrument of the self-indicating or semi-self-indicating type, the prescribed limit of error, in excess or in deficiency, shall be either—

- (a) the appropriate amount specified in the relevant Part of this Schedule for the instrument concerned, or
- (b) the amount corresponding to one half of the smallest interval between consecutive graduations on the scale or dial of the instrument,

whichever is the less.

PART II

BEAM SCALES AND BALANCES

1. *Beam scales marked "Class B".*

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 ounce	#grain	1#grain	# grain	1# grain
8 ounces	1 grain	3 grains	1 grain	2 grains
1 pound	1 grain	3 grains	1 grains	2 grains
2 pounds	1½ grains	4½ grains	2 grains	4 grains
4 pounds	3 grains	9 grains	4 grains	8 grains
7 pounds	4 grains	12 grains	6 grains	12 grains
10 pounds	6 grains	18 grains	9 grains	18 grains
14 pounds	8 grains	24 grains	12 grains	24 grains
28 pounds	15 grains	45 grains	22 grains	44 grains
56 pounds	25 grains	75 grains	40 grains	80 grains
112 pounds	1½ drams	4 ½ drams	2 ½ drams	5 drams
224 pounds	2½ drams	7½ drams	3½ drams	7 drams
Above 2 hundredweight	Add ½ dram for each hundred- weight of capacity	Add 1½ drams for each hundred- weight of capacity	Add 1 dram for each hundred- weight of capacity	Add 2 drams for each hundred- weight of capacity

2. *Beam scales marked “Class C”.*

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 ounce	# grain	1# grain	# grain	1# grains
8 ounces	3 grains	6 grains	3 grains	6 grains
1 pound	3 grains	6 grains	3 grains	6 grains
2 pounds	4½ grains	9 grains	6 grains	12 grains
4 pounds	9 grains	18 grains	12 grains	24 grains
7 pounds	12 grains	24 grains	18 grains	36 grains
10 pounds	18 grains	36 grains	27 grains	54 grains
14 pounds	24 grains	48 grains	36 grains	72 grains

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
28 pounds	45 grains	90 grains	66 grains	132 grains
56 pounds	75 grains	150 grains	120 grains	240 grains
112 pounds	4½ drams	9 drams	7½ drams	15 drams
224 pounds	7½ drams	15 drams	10½ drams	21 drams
Above 2 hundredweight	Add 1½ drams for each hundred- weight of capacity	Add 3 drams for each hundred- weight of capacity	Add 3 drams for each hundred- weight of capacity	Add 6 drams for each hundred- weight of capacity

3. Balances.

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 ounce	120 grain	320 grain	110 grain	# grain
1 pound	110 grain	310 grain	# grain	# grain
7 pounds	½ grain	1½ grains	1 grain	2 grains
56 pounds	1½ grains	4½ grains	2 grains	4 grains

PART III

COUNTER MACHINES

<i>Capacity of machine Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 pound	20 grains	60 grains	30 grains	60 grains
2 pounds	28 grains	84 grains	1½ drams	3 drams
4 pounds	40 grains	120 grains	2 drams	4 drams

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of machine Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitivity when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
7 pounds	2 drams	6 drams	3 drams	6 drams
10 pounds	2½ drams	7½ drams	3½ drams	7 drams
14 pounds	3 drams	9 drams	4½ drams	9 drams
28 pounds	4 drams	12 drams	6 drams	12 drams
56 pounds	6 drams	18 drams	9 drams	18 drams
1 hundredweight	8 drams	24 drams	16 drams	32 drams

PART IV
SPRING BALANCES

<i>Capacity of spring balance Upon passing as fit for use for trade</i>	<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	
1 pound	30 grains	60 grains
2 pounds	1½ drams	3 drams
3 pounds	1½ drams	3 drams
4 pounds	2 drams	4 drams
5 pounds	2½ drams	5 drams
6 pounds	2½ drams	5 drams
7 pounds	3 drams	6 drams
10 pounds	3½ drams	7 drams
11 pounds	3½ drams	7 drams
12 pounds	4 drams	8 drams
13 pounds	4 drams	8 drams
14 pounds	4½ drams	9 drams
15 pounds	4½ drams	9 drams
20 pounds	5 drams	10 drams
21 pounds	5 drams	10 drams
22 pounds	5 drams	10 drams
23 pounds	5½ drams	11 drams
24 pounds	5½ drams	11 drams

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of spring balance Upon passing as fit for use for trade</i>	<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	
25 pounds	5½ drams	11 drams
26 pounds	5½ drams	11 drams
27 pounds	6 drams	12 drams
28 pounds	6 drams	12 drams
29 pounds	6 drams	12 drams
30 pounds	6 drams	12 drams
40 pounds and above	The weight corresponding to ¼ of the interval between consecutive graduations	The weight corresponding to ½ of the interval between consecutive graduations

PART V STEELYARDS

<i>Weight to be added to test sensitivity when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		
<i>Capacity of steelyard</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>
56 pounds	12 drams	36 drams	18 drams	36 drams
1 hundredweight	1 ounce	3 ounces	2 ounces	4 ounces
3 hundredweight	2 ounces	6 ounces	4 ounces	8 ounces
5 hundredweight	3 ounces	9 ounces	6 ounces	12 ounces
7 hundredweight	4 ounces	12 ounces	8 ounces	16 ounces
10 hundredweight	6 ounces	18 ounces	12 ounces	24 ounces
20 hundredweight	10 ounces	30 ounces	20 ounces	40 ounces
30 hundredweights	13 ounces	39 ounces	26 ounces	52 ounces
40 hundredweights	16 ounces	48 ounces	32 ounces	64 ounces
50 hundredweights	20 ounces	60 ounces	40 ounces	80 ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

PART VI
DEAD-WEIGHT MACHINES

<i>Capacity of machine Upon passing as fit for the for trade</i>	<i>Vibrating weighing instruments</i>		<i>Accelerating weighing instruments</i>				<i>Weight required to bring back the beam from position of greatest displacement when fully loaded</i>
	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	
1 hundredweight	½ ounce	1½ ounces	1 ounce	2 ounces	1 ounce	2 ounces	2 ounces
3 hundredweight	1 ounce	3 ounces	2 ounces	4 ounces	2 ounces	4 ounces	4 ounces
5 hundredweight	1½ ounces	4½ ounces	3 ounces	6 ounces	3 ounces	6 ounces	6 ounces
7 hundredweight	2 ounces	6 ounces	4 ounces	8 ounces	4 ounces	8 ounces	8 ounces
10 hundredweight	3 ounces	9 ounces	6 ounces	12 ounces	6 ounces	12 ounces	12 ounces
20 hundredweight	5 ounces	15 ounces	10 ounces	20 ounces	10 ounces	20 ounces	20 ounces
30 hundredweight	6½ ounces	19½ ounces	13 ounces	26 ounces	13 ounces	26 ounces	26 ounces
40 hundredweight	8 ounces	24 ounces	16 ounces	32 ounces	16 ounces	32 ounces	32 ounces
50 hundredweight	10 ounces	30 ounces	20 ounces	40 ounces	20 ounces	40 ounces	40 ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

PART VII

PLATFORM WEIGHING MACHINES

<i>Vibrating weighing instruments</i>		<i>Accelerating weighing instruments</i>				<i>Machines with dials</i>			
<i>Capacity of machine</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>
<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>In relation to the obliteration of stamps</i>
1 hundred-weight	½ ounce	1½ ounces	1 ounce	2 ounces	1 ounce	2 ounces	2 ounces	2 ounces	4 ounces
3 hundred-weight	1 ounce	3 ounces	2 ounces	4 ounces	2 ounces	4 ounces	4 ounces	4 ounces	8 ounces
5 hundred-weight	1½ ounces	4½ ounces	3 ounces	6 ounces	3 ounces	6 ounces	6 ounces	6 ounces	12 ounces
7 hundred-weight	2 ounces	6 ounces	4 ounces	8 ounces	4 ounces	8 ounces	8 ounces	8 ounces	16 ounces
10 hundred-weight	3 ounces	9 ounces	6 ounces	12 ounces	6 ounces	12 ounces	12 ounces	12 ounces	24 ounces
20 hundred-weight	5 ounces	15 ounces	10 ounces	20 ounces	10 ounces	20 ounces	20 ounces	20 ounces	40 ounces
30 hundred-weight	6½ ounces	19½ ounces	13 ounces	26 ounces	13 ounces	26 ounces	26 ounces	26 ounces	52 ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of machine</i>	<i>Error in excess or in deficiency when fully loaded</i>	
<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
12 hundred weight	2 pounds 9 ounces	5 pounds 2 ounces
15 hundred weight	3 pounds 0 ounces	6 pounds 0 ounces
20 hundred weight	3 pounds 12 ounces	7 pounds 8 ounces
30 hundred weight	4 pounds 14 ounces	9 pounds 12 ounces
40 hundred weight	6 pounds 0 ounces	12 pounds 0 ounces
50 hundred weight	7 pounds 8 ounces	15 pounds 0 ounces

PART IX WEIGHBRIDGES

<i>Capacity of machine</i>	<i>Vibrating weighing instruments without dials</i>		<i>Accelerating weighing instruments without dials</i>		<i>Instruments with dials</i>				
	<i>Weight to be added to test when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>				
<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>		
1 ton	1½ pounds	4½ pounds	1½ pounds	3 pounds	1½ pounds	3 pounds	4 pounds	3 pounds	6 pounds
2 tons	2 pounds	6 pounds	2 pounds	4 pounds	2 pounds	4 pounds	5 pounds	4 pounds	8 pounds
5 tons	3½ pounds	10 ½ pounds	4 pounds	8 pounds	4 pounds	8 pounds	10 pounds	8 pounds	16 pounds

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of machine</i>	<i>Vibrating weighing instruments without dials</i>			<i>Accelerating weighing instruments without dials</i>			<i>Instruments with dials</i>		
	<i>Weight to be added to test sensitiveness when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>
	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>
10 tons	5 pounds	15 pounds	6 pounds	12 pounds	6 pounds	12 pounds	15 pounds	12 pounds	24 pounds
20 tons	7 pounds	21 pounds	10 pounds	20 pounds	10 pounds	20 pounds	25 pounds	20 pounds	40 pounds
25 tons	8 pounds	24 pounds	12 pounds	24 pounds	12 pounds	24 pounds	30 pounds	24 pounds	48 pounds
30 tons	8½ pounds	25½ pounds	13½ pounds	27 pounds	13½ pounds	27 pounds	34 pounds	27 pounds	54 pounds
35 tons	9 pounds	27 pounds	15 pounds	30 pounds	15 pounds	30 pounds	37 pounds	30 pounds	60 pounds
40 tons	9½ pounds	28½ pounds	16 pounds	32 pounds	16 pounds	32 pounds	40 pounds	32 pounds	64 pounds
50 tons	10 pounds	30 pounds	18 pounds	36 pounds	18 pounds	36 pounds	45 pounds	36 pounds	52 pounds
75 tons	12 pounds	36 pounds	23 pounds	46 pounds	23 pounds	46 pounds	58 pounds	46 pounds	92 pounds
100 tons	14 pounds	42 pounds	28 pounds	56 pounds	28 pounds	56 pounds	70 pounds	56 pounds	112 pounds
200 tons	18 pounds	54 pounds	42 pounds	84 pounds	42 pounds	84 pounds	105 pounds	84 pounds	168 pounds

PART X

CRANE WEIGHING MACHINES

Crane weighing machines constructed upon other than the hydraulic principle

<i>Capacity of machine Upon passing as fit for use for trade</i>	<i>Machines with steelyard indicators</i>		<i>Machines with dials</i>			
	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>In relation to the obliteration of stamps</i>
1 hundredweight	½ ounce	1½ ounces	1 ounces	2 ounces	2 ounces	4 ounces
5 hundredweight	1½ ounces	4½ ounces	3 ounces	6 ounces	6 ounces	12 ounces
10 hundredweight	3 ounces	9 ounces	6 ounces	12 ounces	12 ounces	1½ pounds
1 ton	1½ pounds	4½ pounds	1½ pounds	3 pounds	3 pounds	6 pounds
2 tons	2 pounds	6 pounds	2 pounds	4 pounds	4 pounds	8 pounds
5 tons	3½ pounds	10½ pounds	4 pounds	8 pounds	8 pounds	16 pounds
10 tons	5 pounds	15 pounds	6 pounds	12 pounds	12 pounds	24 pounds
20 tons	7 pounds	21 pounds	10 pounds	20 pounds	20 pounds	40 pounds
25 tons	8 pounds	24 pounds	12 pounds	24 pounds	24 pounds	48 pounds
30 tons	8½ pounds	25½ pounds	13½ pounds	27 pounds	27 pounds	54 pounds
35 tons	9 pounds	27 pounds	15 pounds	30 pounds	30 pounds	60 pounds
40 tons	9½ pounds	28½ pounds	16 pounds	32 pounds	32 pounds	64 pounds
50 tons	10 pounds	30 pounds	18 pounds	36 pounds	36 pounds	72 pounds
75 tons	12 pounds	36 pounds	23 pounds	46 pounds	46 pounds	92 pounds
100 tons	14 pounds	42 pounds	28 pounds	56 pounds	56 pounds	112 pounds
200 tons	18 pounds	54 pounds	42 pounds	84 pounds	84 pounds	168 pounds

Crane weighing machines constructed upon the hydraulic principle

2. An amount equal to one-half of the weight represented by the interval between consecutive graduations.

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

PART XI

AUTOMATIC WEIGHING MACHINES

<i>Error in relation to the passing as fit for use for trade</i>		<i>Error in relation to the obliteration of stamps</i>					
<i>For the purposes of Test A</i>		<i>For the purposes of Test A</i>		<i>For the purposes of Test B</i>		<i>For the purposes of Test B</i>	
<i>Description of machine</i>	<i>Capacity of machine</i>	<i>When tested by means of test loads</i>	<i>When tested by means of the direct application of appropriate weights</i>	<i>For the purposes of Test B</i>	<i>When tested by means of test loads</i>	<i>When tested by means of the direct application of appropriate weights</i>	<i>For the purposes of Test B</i>
<i>Column 1</i>	<i>Column 2</i>	<i>Column 3</i>	<i>Column 4</i>	<i>Column 5</i>	<i>Column 6</i>	<i>Column 7</i>	<i>Column 8</i>
All automatic weighing machines, other than those hereinafter described in columns 1 and 2 of this table	Under 10 pounds	Under ½ per cent., in excess only, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	½ per cent., in excess only, of the purported weight of the test load	1 per cent., in excess only, of the purported weight of the test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	1 per cent., in excess only, of the purported weight of the test load
10 pounds or more	½ per cent., in excess only, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	1 per cent., in excess or in deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	Test not applicable

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Error in relation to the passing as fit for use for trade</i>		<i>Error in relation to the obliteration of stamps</i>					
<i>For the purposes of Test A</i>		<i>For the purposes of Test A</i>				<i>For the purposes of Test B</i>	
<i>Description of machine</i>	<i>Capacity of machine</i>	<i>When tested by means of test loads</i>	<i>When tested by means of appropriate weights</i>	<i>For the purposes of Test B</i>	<i>When tested by means of test loads</i>	<i>When tested by means of appropriate weights</i>	<i>For the purposes of Test B</i>
<i>Column 1</i>	<i>Column 2</i>	<i>Column 3</i>	<i>Column 4</i>	<i>Column 5</i>	<i>Column 6</i>	<i>Column 7</i>	<i>Column 8</i>
Automatic weighing machines for use only for weighing grain	10 pounds or more	¼ per cent., in excess or deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	½ per cent., in excess or deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of weighing instrument to which the machine most nearly relates	Test not applicable
Automatic weighing machines for use only for weighing solid fuel	2 hundredweight or less	2 per cent., in excess or deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	2 per cent., in excess only, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable
Automatic weighing machines for use	Any capacity	20 per cent., in excess only, of	Prescribed limit of error applicable	Test not applicable	20 per cent., in excess only, of	Prescribed limit of error applicable	Test not applicable

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Error in relation to the passing as fit for use for trade</i>		<i>Error in relation to the obliteration of stamps</i>					
<i>For the purposes of Test A</i>		<i>For the purposes of Test A</i>		<i>For the purposes of Test B</i>		<i>For the purposes of Test B</i>	
<i>Description of machine</i>	<i>Capacity of machine</i>	<i>When tested by means of test loads</i>	<i>When tested by means of the direct application of appropriate weights</i>	<i>For the purposes of Test B</i>	<i>When tested by means of test loads</i>	<i>When tested by means of the direct application of appropriate weights</i>	<i>For the purposes of Test B</i>
<i>Column 1</i>	<i>Column 2</i>	<i>Column 3</i>	<i>Column 4</i>	<i>Column 5</i>	<i>Column 6</i>	<i>Column 7</i>	<i>Column 8</i>
only for weighing potato crisps		the total purported weight of 20 test loads	to the type, class or description of the weighing instrument to which the machine most nearly relates		the total purported weight of 20 test loads	to the type, class or description of weighing instrument to which the machine most nearly relates	
Totalising weighing machines,	Any capacity	½ per cent., in excess or in deficiency, of the total purported weight of 40 test loads	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	1 per cent., in excess or in deficiency, of the total purported weight of 40 test loads	Prescribed limit of error applicable to the type, class or description of weighing instrument to which the machine most nearly relates	Test not applicable

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

SCHEDULE 3

PRINCIPLES OF CONSTRUCTION OF AVOIRDUPOIS WEIGHTS OF THE BELL TYPE

Table Purported values and heights

<i>Column 1</i>		<i>Column 2</i>	
<i>(a)</i>		<i>(b)</i>	
	<i>Purported value of weight</i>	<i>Height of weight</i>	<i>Limits</i>
inches		inches	
	¼ ounce	¾	116
	½ ounce	1	116
		132	
	1 ounce	1¼	116
	2 ounces	1⅝	116
	4 ounces	2	116
		116	
	8 ounces	2⅜	116
	1 pound	2⅞	⅛
	2 pounds	3⅜	⅛
	4 pounds	4¼	⅛
	7 pounds	5	⅛
		116	
	14 pounds	7⅞	⅛

Diagram