
STATUTORY INSTRUMENTS

1995 No. 1907

The Non-automatic Weighing Instruments
(EEC Requirements) Regulations 1995

PART III

USE FOR TRADE OF NON-AUTOMATIC WEIGHING INSTRUMENTS

Instruments marked with temperature range

30. Where an instrument is marked with a temperature range, a person shall not use the instrument for trade at temperatures outside that range.