
STATUTORY INSTRUMENTS

1995 No. 735

**The Measuring Equipment (Capacity Measures
and Testing Equipment) Regulations 1995**

PART V

TESTING EQUIPMENT—PERISTALTIC PUMP DISPENSERS

Peristaltic pump dispensers

19. The Weights and Measures (Local and Working Standard Capacity Measures and Testing Equipment) Regulations 1990(1) are hereby amended,—

- (a) in regulation 2(1), by the insertion after the definition of “multifiller” of the following definition—

““peristaltic pump dispenser” means a device, operating on the peristaltic principle, designed to deliver a known quantity of water for the testing of dispensing measures for pharmaceutical purposes;”;

- (b) in Part III, by the insertion after regulation 18 of the following regulation—

“Peristaltic pump dispensers

18A.—(1) A peristaltic pump dispenser shall have been tested within 6 months before use at the minimum, at the maximum and at five intermediate delivered volumes suitable for the purpose for which it is to be used.

(2) A peristaltic pump dispenser shall be tested by weighing the water delivered by a suitable weighing machine and determining the delivered volume from the weight so ascertained.

(3) The error on a peristaltic pump dispenser shall not exceed the limits of error shown in Schedule 2 to these Regulations.”; and

- (c) in Schedule 2, by the insertion at the end of the following table—

“Peristaltic pump dispensers

Delivered volume millilitres	Limit of error millilitres
1	0.04
2	0.04
3	0.05
4	0.05

Changes to legislation: There are currently no known outstanding effects for the The Measuring Equipment (Capacity Measures and Testing Equipment) Regulations 1995, PART V. (See end of Document for details)

Delivered volume millilitres	Limit of error millilitres
5	0.06
6	0.07
7	0.07
8	0.08
9	0.08
10	0.08
15	0.10
20	0.12
25	0.12
30	0.13
35	0.13
40	0.14
45	0.14
50	0.15
60	0.16
70	0.17
75	0.17
80	0.18
90	0.20
100	0.20
125	0.25
150	0.30
175	0.35
200	0.40
225	0.40
250	0.40
300	0.42
350	0.42
400	0.45
450	0.45
500	0.50
1000	1.00”.

Changes to legislation: There are currently no known outstanding effects for the The Measuring Equipment (Capacity Measures and Testing Equipment) Regulations 1995, PART V. (See end of Document for details)

.....

Commencement Information

II [Reg. 19](#) in force at 10.4.1995, see [reg. 1\(1\)](#)

Changes to legislation:

There are currently no known outstanding effects for the The Measuring Equipment (Capacity Measures and Testing Equipment) Regulations 1995, PART V.