
STATUTORY INSTRUMENTS

1995 No. 735

**The Measuring Equipment (Capacity Measures
and Testing Equipment) Regulations 1995**

PART III

TESTING

Testing of nominal capacity and scale marks of capacity measures

10.—(1) The inspector shall test the nominal capacity and all the scale marks or, in the case of a graduated measure, such of them as he considers necessary.

(2) The capacity of a measure shall be the volume at 20 °C when the measure is on a horizontal surface and—

- (a) in the case of a line measure, the bottom part of the meniscus coincides with the top of the line, or
- (b) in the case of a brim measure, the surface of the liquid is in the same plane as the brim.

Commencement Information

II [Reg. 10](#) in force at 10.4.1995, see [reg. 1\(1\)](#)

Changes to legislation:

There are currently no known outstanding effects for the The Measuring Equipment (Capacity Measures and Testing Equipment) Regulations 1995, Section 10.