
STATUTORY INSTRUMENTS

2006 No. 1268

The Measuring Instruments (Cold-water Meters) Regulations 2006

PART II

PLACING ON THE MARKET AND PUTTING INTO USE OF COLD-WATER METERS

Conformity assessment procedures

- 6.—(1) The conformity assessment procedures are the procedures as follows—
- (a) B and D;
 - (b) B and F; or
 - (c) H1.
- (2) The manufacturer or his authorised representative shall provide to the notified body carrying out the relevant conformity assessment procedure the technical documentation set out in Schedule 3.
- (3) In this regulation—
- (a) “B” means type examination, set out in Annex B;
 - (b) “D” means declaration of conformity to type based on quality assurance of the production process, set out in Annex D;
 - (c) “F” means declaration of conformity to type based on product verification, set out in Annex F; and
 - (d) “H1” means declaration of conformity based on full quality assurance plus design examination, set out in Annex H1.