#### STATUTORY INSTRUMENTS

### 2006 No. 1270

#### WEIGHTS AND MEASURES

# The Measuring Instruments (Non-Prescribed Instruments) Regulations 2006 (revoked)<sup>F1</sup>

Made - - - 28th April 2006

Laid before Parliament 8th May 2006

Coming into force

Regulations 1, 2, 4 6 and 7 and Part 1 of

Schedule 2 30th May 2006

Remaining regulations 30th October 2006

## THE MEASURING INSTRUMENTS (NON-PRESCRIBED INSTRUMENTS) REGULATIONS 2006 (REVOKED)

- 1. Citation and commencement
- 2. Interpretation
- 3. Application
- 4. Designation of notified bodies
- 5. Functions of notified bodies
- 6. Provisions supplemental to regulation 4
- 7. Variation and termination of designations
- 8. Fees
- 9. Compliance with the essential requirements
- 10. Conformity assessment procedures
- 11. Marking and identification requirements
- 12. Conformity with other directives Signature

#### SCHEDULE 1 —

PART 1 — ESSENTIAL REQUIREMENTS

- 1. The essential requirements are the relevant requirements relating to measuring...
- 2. Definitions

- 3. Allowable Errors
- 4. Reproducibility
- 5. Repeatability
- 6. Discrimination and Sensitivity
- 7. Durability
- 8. Reliability
- 9. Suitability
- 10. Protection against corruption
- 11. Information to be borne by and to accompany the instrument
- 12. Indication of result
- 13. Further processing of data to conclude the trading transaction
- 14. Conformity evaluation

#### PART 2 — INSTRUMENT SPECIFIC REQUIREMENTS

#### Part A

#### Water Meters

- 15. The relevant requirements of Part 1 and the specific requirements...
- 16. Definitions
- 17. Rated Operating Conditions
- 18. MPE
- 19. Permissible Effect of Disturbances
- 20. Suitability
- 21. Units of Measurement
- 22. Putting into Use

#### Part B

#### **Heat Meters**

- 23. The relevant requirements of Part 1 and the specific requirements...
- 24. Definitions
- 25. Rated operating conditions
- 26. Accuracy classes
- 27. MPE applicable to complete heat meters
- 28. Permissible influences of electromagnetic disturbances
- 29. Durability
- 30. Inscriptions on a heat meter
- 31. Sub-assemblies
- 32. Putting into use

#### Part C

Measuring System for the Continuous and Dynamic Measurement of Quantities of Liquids other than Water

- 33. (1) The relevant requirements of Part 1 and the specific...
- 34. Definitions
- 35. Rated operating conditions
- 36. Accuracy classification and maximum permissible errors (MPEs)
- 37. Maximum permissible effect of disturbances
- 38. Durability

Changes to legislation: There are currently no known outstanding effects for the The Measuring Instruments (Non-Prescribed Instruments) Regulations 2006 (revoked). (See end of Document for details)

- 39. Suitability
- 40. Power supply failure
- 41. Putting into use
- 42. Units of measurement

#### Part D

#### **Automatic Weighing Instruments**

- 43. The relevant requirements of Part 1 and the specific requirements...
- 44. Rated Operating Conditions
- 45. Suitability
- 46. Automatic Catchweighers
- 47. Automatic gravimetric filling instruments
- 48. Discontinuous Totalisers
- 49. Continuous Totalisers
- 50. Automatic Rail-weighbridges

#### Part E

#### Material Measures

- 51. (1) The relevant essential requirements of Part 1 and the...
- 52. Material measures of length
- 53. Capacity serving measures

#### Part F

#### **Dimensional Measuring Instruments**

- 54. The relevant essential requirements of Part 1 and the specific...
- 55. Requirements common to all dimensional measuring instruments
- 56. Length measuring instruments
- 57. Area measuring instruments
- 58. Instrument The MPE is 1.0 % but not less than...
- 59. Other requirements
- 60. Multi-dimensional measuring instruments

#### Part G

#### Volume Conversion Devices

- 61. The relevant essential requirements of Part 1 and the specific...
- 62. A volume conversion device constitutes a sub-assembly for the purpose...
- 63. Rated operating conditions
- 64. MPE
- 65. Permissible effect of disturbances
- 66. Suitability
- 67. Units
- 68. Base conditions for converted quantities
- 69. Putting into Use

#### SCHEDULE 2 — NOTIFIED BODIES

#### PART 1 — NOTIFIED BODY CRITERIA

- 1. The body, its director and staff involved in conformity assessment...
- 2. The body, its director and staff involved in conformity assessment...
- 3. The conformity assessment shall be carried out with the highest...
- 4. The body shall be capable of carrying out all the...
- 5. The body's staff shall have—(a) sound technical and vocational...
- 6. The impartiality of the body, its director and staff shall...
- 7. The body shall satisfy the designating authority that it has...
- 8. The body's director and staff shall be bound to observe...

#### PART 2 — FUNCTIONS

- 9. Assessment of applications for certificates or notifications
- 10. Limitations on duties to exercise functions
- 11. Contractors
- 12. Form of certificates or notifications
- 13. Conditions of certificates or notifications
- 14. Withdrawal of certificates or notifications
- 15. Procedure where a notified body is minded to refuse to give, or to vary or withdraw a certificate or approval

#### SCHEDULE 3 — TECHNICAL DOCUMENTATION

- 1. The technical documentation shall render the design, manufacture and operation...
- 2. The technical documentation shall be sufficiently detailed to ensure—
- The technical documentation shall include insofar as relevant for assessment
- 4. The manufacturer shall specify where seals and markings have been...
- 5. The manufacturer shall indicate the conditions for compatibility with interfaces...

#### SCHEDULE 4 — MARKING AND INSCRIPTIONS

- 1. The CE marking consists of the symbol "CE" according to...
- 2. The M marking consists of the capital letter "M" and...
- 3. The identification number of the notified body concerned shall follow...
- 4. When a measuring instrument consists of a set of devices,...
- 5. When a measuring instrument is too small or too sensitive...
- 6. The CE marking and the M marking shall be indelible....

**Explanatory Note** 

Changes to legislation:
There are currently no known outstanding effects for the The Measuring Instruments (Non-Prescribed Instruments) Regulations 2006 (revoked).