

## SCHEDULE 1

### PART 2

#### INSTRUMENT SPECIFIC REQUIREMENTS

##### Part C

##### Measuring System for the Continuous and Dynamic Measurement of Quantities of Liquids other than Water

##### **Durability**

**38.** After an appropriate test, taking into account the period of time estimated by the manufacturer, has been performed, the following criterion shall be satisfied—

The variation of the measurement result after the durability test, when compared with the initial measurement result, shall not exceed the value for meters specified in line B of Table 3.