

## SCHEDULE 4

Regulation 11

### MARKING AND INSCRIPTIONS

1. The CE marking consists of the symbol “CE” according to the design laid down in paragraph I.B(d) of the Annex to Decision 93/465/EEC<sup>M1</sup>. The CE marking shall be at least 5 mm high.

#### **Marginal Citations**

**M1** OJ L220, 30.8.93, p.23.

2. The M marking consists of the capital letter “M” and the last two digits of the year of its affixing, surrounded by a rectangle. The height of the rectangle shall be equal to the height of the CE marking. The M marking shall immediately follow the CE marking.

3. The identification number of the notified body concerned shall follow the CE marking and the M marking.

4. When a measuring instrument consists of a set of devices, not being sub-assemblies, operating together, the markings shall be affixed on the instrument's main device.

5. When a measuring instrument is too small or too sensitive to carry the CE marking and the M marking, the markings shall be carried by the packaging, if any, and by the accompanying documents required by these Regulations.

6. The CE marking and the M marking shall be indelible. The identification number of the notified body concerned shall be indelible or self destructive upon removal. All markings shall be clearly visible or easily accessible.

**Status:**

Point in time view as at 30/10/2006.

**Changes to legislation:**

There are currently no known outstanding effects for the The Measuring Instruments (Non-Prescribed Instruments) Regulations 2006 (revoked), SCHEDULE 4.