

SCHEDULE 4

Regulation 12(3)

MARKING AND INSCRIPTIONS

1. The CE marking consists of the symbol “CE” according to the design laid down in paragraph 1.B(d) of the Annex to Decision [93/465/EEC](#)(1). The CE marking shall be at least 5 mm high.
2. The M marking consists of the capital letter “M” and the last two digits of the year of its affixing, surrounded by a rectangle. The height of the rectangle shall be equal to the height of the CE marking. The M marking shall immediately follow the CE marking.
3. The identification number of the notified body concerned shall follow the CE marking and the M marking.
4. When a relevant instrument consists of a set of devices operating together, the markings shall be affixed on the instrument’s main device.
5. The CE marking and the M marking shall be indelible. The identification number of the notified body concerned shall be indelible or self-destructive upon removal. All markings shall be clearly visible or easily accessible.

(1) O.J. No. L220, 30.8.93, p.23.