

SCHEDULE 4 **U.K.**

Regulation 12(3)

MARKING AND INSCRIPTIONS

1. The CE marking consists of the symbol “CE” according to the design laid down in paragraph 1.B(d) of the Annex to Decision [93/465/EEC\(1\)](#). The CE marking shall be at least 5 mm high.

Commencement Information

I1 Sch. 4 para. 1 in force at 30.10.2006, see [reg. 1\(2\)](#)

2. The M marking consists of the capital letter “M” and the last two digits of the year of its affixing, surrounded by a rectangle. The height of the rectangle shall be equal to the height of the CE marking. The M marking shall immediately follow the CE marking.

Commencement Information

I2 Sch. 4 para. 2 in force at 30.10.2006, see [reg. 1\(2\)](#)

3. The identification number of the notified body concerned shall follow the CE marking and the M marking.

Commencement Information

I3 Sch. 4 para. 3 in force at 30.10.2006, see [reg. 1\(2\)](#)

4. When a relevant instrument consists of a set of devices operating together, the markings shall be affixed on the instrument’s main device.

Commencement Information

I4 Sch. 4 para. 4 in force at 30.10.2006, see [reg. 1\(2\)](#)

5. The CE marking and the M marking shall be indelible. The identification number of the notified body concerned shall be indelible or self-destructive upon removal. All markings shall be clearly visible or easily accessible.

Commencement Information

I5 Sch. 4 para. 5 in force at 30.10.2006, see [reg. 1\(2\)](#)

(1) O.J. No. L220, 30.8.93, p.23.

Changes to legislation:

There are outstanding changes not yet made by the legislation.gov.uk editorial team to The Measuring Instruments (Gas Meters) Regulations 2006. Any changes that have already been made by the team appear in the content and are referenced with annotations.

[View outstanding changes](#)

Changes and effects yet to be applied to :

- Regulations revoked by [S.I. 2016/1153 Sch. 3 para. 1 table](#)

Changes and effects yet to be applied to the whole Instrument associated Parts and Chapters:

Whole provisions yet to be inserted into this Instrument (including any effects on those provisions):

- Sch. 1 para. 13(5) inserted by [S.I. 2010/2881 reg. 7\(c\)\(ii\)](#)