

## SCHEDULE 2

### Amendment of the Medical Devices (Amendment etc.) (EU Exit) Regulations 2019

#### **Insertion of regulation 4(6B)**

15. After regulation 4(6A) insert—

“(6B) For regulation 11 (CE marking of general medical devices that come within the scope of more than one Directive) substitute—

#### **“UK marking of general medical devices that come within the scope of this Part and other legislation**

11. Where a relevant device (within the meaning of this Part) comes within the scope of this Part and other product safety or health and safety legislation (“the other legislation”) a person must not affix a UK marking to the device unless the relevant requirements of the other legislation are also satisfied.”.