

SCHEDULE 2

Amendment of the Medical Devices (Amendment etc.) (EU Exit) Regulations 2019

Insertion of regulation 6(4B)

39. After regulation 6(4A) insert—

“(4B) For regulation 37 (CE marking of *in vitro* diagnostic medical devices that come within the scope of more than one Directive) substitute—

“UK marking of *in vitro* diagnostic devices that come within the scope of this Part and other legislation

37. Where a relevant device (within the meaning of this Part) comes within the scope of this Part and other product safety or health and safety legislation (“the other legislation”) a person must not affix a UK marking to the device unless the relevant requirements of the other legislation are also satisfied.”.