
STATUTORY INSTRUMENTS

2021 No. 905

The Medical Devices (Northern
Ireland Protocol) Regulations 2021

PART 8

Amendment of the Medical Devices Regulations 2002

Amendment of regulation 10A (UK(NI) indication: general medical devices)

33. In regulation 10A, after paragraph (3), insert—

“(3A) The UK(NI) indication may be less than 5mm high provided that it is the same height as the CE marking that it accompanies.”