

SCHEDULE 5

SAMPLING

Dynamic spear sampler

14. In the table set out in paragraph 13 references to type A, B, C and D instruments shall be construed as referring to instruments having a point length (a), shoulder length (b), boss length (c), aperture length (d), aperture width (e), bore width (f) and outside diameter (g) as shown in the diagram below of the dimensions specified in the corresponding entries in the following table—

<i>Type of Instrument</i>	<i>Point length mm (a)</i>	<i>Shoulder length mm (b)</i>	<i>Boss length mm (c)</i>	<i>Aperture length mm (d)</i>	<i>Aperture width mm (e)</i>	<i>Bore width mm (f)</i>	<i>Outside diameter mm (g)</i>
A	42	7	8	20	8	10	12
B	85	12	10	33	11	13	15
C	82	12	13	40	15	17	19
D	78	15	15	40	18	20	22

